

Blood Pressure Service (BLS)

Bluetooth® Implementation Conformance Statement (ICS) Proforma

- **Revision:** BLS.ICS.p8
- **Revision Date:** 2024-07-01
- **Prepared By:** Medical Devices Working Group
- **Published during TCRL:** TCRL.2024-1



This document, regardless of its title or content, is not a Bluetooth Specification as defined in the Bluetooth Patent/Copyright License Agreement (“PCLA”) and Bluetooth Trademark License Agreement. Use of this document by members of Bluetooth SIG is governed by the membership and other related agreements between Bluetooth SIG Inc. (“Bluetooth SIG”) and its members, including the PCLA and other agreements posted on Bluetooth SIG’s website located at www.bluetooth.com.

THIS DOCUMENT IS PROVIDED “AS IS” AND BLUETOOTH SIG, ITS MEMBERS, AND THEIR AFFILIATES MAKE NO REPRESENTATIONS OR WARRANTIES AND DISCLAIM ALL WARRANTIES, EXPRESS OR IMPLIED, INCLUDING ANY WARRANTY OF MERCHANTABILITY, TITLE, NON-INFRINGEMENT, FITNESS FOR ANY PARTICULAR PURPOSE, THAT THE CONTENT OF THIS DOCUMENT IS FREE OF ERRORS.

TO THE EXTENT NOT PROHIBITED BY LAW, BLUETOOTH SIG, ITS MEMBERS, AND THEIR AFFILIATES DISCLAIM ALL LIABILITY ARISING OUT OF OR RELATING TO USE OF THIS DOCUMENT AND ANY INFORMATION CONTAINED IN THIS DOCUMENT, INCLUDING LOST REVENUE, PROFITS, DATA OR PROGRAMS, OR BUSINESS INTERRUPTION, OR FOR SPECIAL, INDIRECT, CONSEQUENTIAL, INCIDENTAL OR PUNITIVE DAMAGES, HOWEVER CAUSED AND REGARDLESS OF THE THEORY OF LIABILITY, AND EVEN IF BLUETOOTH SIG, ITS MEMBERS, OR THEIR AFFILIATES HAVE BEEN ADVISED OF THE POSSIBILITY OF SUCH DAMAGES.

This document is proprietary to Bluetooth SIG. This document may contain or cover subject matter that is intellectual property of Bluetooth SIG and its members. The furnishing of this document does not grant any license to any intellectual property of Bluetooth SIG or its members.

This document is subject to change without notice.

Copyright © 2011–2024 by Bluetooth SIG, Inc. The Bluetooth word mark and logos are owned by Bluetooth SIG, Inc. Other third-party brands and names are the property of their respective owners.



Contents

1	Identification of the implementation.....	4
1.1	Implementation Under Test (IUT) identification	4
1.2	Versions	5
1.3	Transports	5
1.4	Service requirements	5
1.5	GATT requirements	8
1.6	GAP requirements	8
2	References.....	9
3	Revision history and acknowledgments	10

1 Identification of the implementation

1.1 Implementation Under Test (IUT) identification

Identification of the Implementation Under Test (IUT) is to be filled in to provide as much detail as possible regarding version numbers and configuration options.

An ICS contact person to respond to queries regarding information supplied in this ICS proforma is named in the Declaration of Compliance: Summary of Selected Specifications in Implementation.

1.2 Versions

Table 0: X.Y Versions

Item	Version	Reference	Status
1	BLS v1.0	[1]	Deprecated 2023-02-01. Withdrawn 2024-02-01.
2	BLS v1.1	[4]	C.3, C.4

C.1–C.2: No longer used.

C.3: Mandatory.

C.4: Can only be supported with an active X.Y.Z version after Deprecation. Deprecated 2023-02-01.
Withdrawn 2025-02-01.

Table 0b: X.Y.Z Versions

Item	Version	Reference	Status
1	BLS v1.1.1	[4]	C.1

C.1: Mandatory IF BLS 0/2 “BLS v1.1”, otherwise Excluded.

1.3 Transports

Table 1: Transport Requirements

Item	Transport	Reference	Status
1	Service supported over BR/EDR	[1] 1.5	C.1
2	Service supported over LE	[1] 1.5	M

C.1: Excluded for this Service.

1.4 Service requirements

Table 2: Service Requirements

Item	Capability	Reference	Status
1	Blood Pressure Service	[1] 2	M
2	Blood Pressure Measurement Characteristic	[1] 3.1	M
3	Time Stamp field of Blood Pressure Measurement Characteristic	[1] 3.1.1.3	C.1
4	Pulse Rate field of Blood Pressure Measurement Characteristic	[1] 3.1.1.4	O
5	User ID field of Blood Pressure Measurement Characteristic	[1] 3.1.1.5	C.2
6	Measurement Status field of Blood Pressure Measurement Characteristic	[1] 3.1.1.6	O
7	Intermediate Cuff Pressure Characteristic	[1] 3.2	O
8	Time Stamp field of Intermediate Cuff Pressure Characteristic	[1] 3.2.1.3	C.4

Item	Capability	Reference	Status
9	Pulse Rate field of Intermediate Cuff Pressure Characteristic	[1] 3.2.1.4	C.4
10	User ID field of Intermediate Cuff Pressure Characteristic	[1] 3.2.1.5	C.3
11	Measurement Status field of Intermediate Cuff Pressure Characteristic	[1] 3.2.1.6	C.4
12	Blood Pressure Feature Characteristic	[1] 3.3	M
13	Stored Measurements	[1] 3.4 [4] 3.8	C.5
14	Multi-user Support	[1] 3.1.1.5, 3.2.1.5 [4] 3.4, 3.5	O
15	Multiple Bond	[1] 3.3.1 [4] 3.3.1.6	O
16	Enhanced Blood Pressure Service	[4] 3	O
17	E2E-CRC	[4] 3.3.1.7	C.7
18	User Data Service	[4] 3.3.1.8	C.8
19	User Facing Time	[4] 3.3.1.9	C.8
20	Enhanced Blood Pressure Measurement Characteristic	[4] 3.4	C.8, C.9
21	Enhanced Intermediate Cuff Pressure Characteristic	[4] 3.5	O
22	RACP Characteristic	[4] 3.6	C.8
23	Blood Pressure Record Characteristic	[4] 3.7	C.10
24	Epoch Start 2000 Flag	[4] 3.4.1.1, 3.5.1.1	O
25	Indications for changes of supported features	[5] 3.3.1	M

C.1: Mandatory IF BLS 2/13 “Stored Measurements”, otherwise Optional.

C.2: Mandatory IF BLS 2/14 “Multi-user Support”, otherwise Optional.

C.3: Mandatory IF BLS 2/7 “Intermediate Cuff Pressure Characteristic” AND BLS 2/14 “Multi-user Support”, otherwise Optional.

C.4: Optional IF BLS 2/7 “Intermediate Cuff Pressure Characteristic”, otherwise Excluded.

C.5: Mandatory IF BLS 2/22 “RACP Characteristic”, otherwise Optional.

C.6: No longer used.

C.7: Optional IF BLS 2/23 “Blood Pressure Record Characteristic”, otherwise Excluded.

C.8: Mandatory to support at least one IF BLS 2/16 “Enhanced Blood Pressure Service”, otherwise Excluded.

C.9: Mandatory IF BLS 2/19 “User Facing Time”, otherwise Optional.

C.10: Mandatory IF BLS 2/22 “RACP Characteristic”, otherwise Excluded.

Table 2a: Service Requirements – Enhanced Blood Pressure Measurement

Prerequisite: BLS 2/20 “Enhanced Blood Pressure Measurement Characteristic”

Item	Capability	Reference	Status
1	Time Stamp field of Enhanced Blood Pressure Measurement Characteristic	[4] 3.4, 3.4.1.3	C.1

Item	Capability	Reference	Status
2	Pulse Rate field of Enhanced Blood Pressure Measurement Characteristic	[4] 3.4, 3.4.1.4	O
3	User ID field of Enhanced Blood Pressure Measurement Characteristic	[4] 3.4, 3.4.1.5	C.2
4	Measurement Status field of Enhanced Blood Pressure Measurement Characteristic	[4] 3.4, 3.4.1.6	O
5	User Facing Time field of Enhanced Blood Pressure Measurement Characteristic	[4] 3.4, 3.4.1.7	C.3

C.1: Mandatory IF BLS 2/13 “Stored Measurements” OR BLS 2/22 “RACP Characteristic”, otherwise Optional.

C.2: Mandatory IF BLS 2/14 “Multi-user Support”, otherwise Optional.

C.3: Optional IF BLS 2/19 “User Facing Time”, otherwise Excluded.

Table 2b: Service Requirements – Enhanced Intermediate Cuff Pressure

Prerequisite: BLS 2/21 “Enhanced Intermediate Cuff Pressure Characteristic”

Item	Capability	Reference	Status
1	Time Stamp field of Enhanced Intermediate Cuff Pressure Characteristic	[4] 3.5, 3.5.1.3	O
2	Pulse Rate field of Enhanced Intermediate Cuff Pressure Characteristic	[4] 3.5, 3.5.1.4	O
3	User ID field of Enhanced Intermediate Cuff Pressure Characteristic	[4] 3.5, 3.5.1.5	C.1
4	Measurement Status field of Enhanced Intermediate Cuff Pressure Characteristic	[4] 3.5, 3.5.1.6	O
5	User Facing Time field of Enhanced Intermediate Cuff Pressure Characteristic	[4] 3.5, 3.5.1.7	C.2

C.1: Mandatory IF BLS 2/14 “Multi-user Support”, otherwise Excluded.

C.2: Optional IF BLS 2/19 “User Facing Time”, otherwise Excluded.

Table 2c: Service Requirements – Record Access Control Point

Prerequisite: BLS 2/22 “RACP Characteristic”

Item	Capability	Reference	Status
1	Report Stored Records	[4] 3.6.2.1, 3.6.3.1	M
2	Report Number of Stored Records	[4] 3.6.2.1, 3.6.3.2	M
3	Delete Stored Records	[4] 3.6.2.1, 3.6.3.3	O
4	Abort Operation	[4] 3.6.2.1, 3.6.3.4	O

Table 2d: Indications for changes of supported features

Prerequisite: BLS 2/25 “Indications for changes of supported features”

Item	Capability	Reference	Status
1	Changeable Blood Pressure Feature	[5] 3, 3.3.1	O

Item	Capability	Reference	Status
2	Blood Pressure Feature characteristic indication	[5] 3, 3.3.1	C.1

C.1: Mandatory IF BLS 2d/1 “Changeable Blood Pressure Feature” AND BLS 4/2 “Bonding procedure”, otherwise Excluded.

1.5 GATT requirements

Table 3: GATT Requirements

Item	Capability	Reference	Status	Inter-Layer Dependency
1	No longer used	N/A	N/A	N/A
2	GATT Server over LE	[1] 1.5	M	[2] GATT 1a/3
3	Notifications	[1] 1.4	C.1	[2] GATT 4/17
4	Indications	[1] 1.4	M	[2] GATT 4/18
5	Read Characteristic Descriptors	[1] 1.4	M	[2] GATT 4/19
6	Write Characteristic Descriptors	[1] 1.4	M	[2] GATT 4/21

C.1: Mandatory IF BLS 2/7 “Intermediate Cuff Pressure Characteristic” OR BLS 2/21 “Enhanced Intermediate Cuff Pressure Characteristic” OR BLS 2/23 “Blood Pressure Record Characteristic”, otherwise not defined.

1.6 GAP requirements

Table 4: GAP Requirements

Item	Capability	Reference	Status	Inter-Layer Dependency
1	Bondable mode	[1] 3	C.1	[3] GAP 24/2
2	Bonding procedure	[1] 3	O	[3] GAP 24/3

C.1: Mandatory IF BLS 4/2 “Bonding procedure”, otherwise not defined.

2 References

- [1] Blood Pressure Service Specification, Version 1.0 or later
- [2] ICS Proforma for Generic Attribute Profile (GATT)
- [3] ICS Proforma for Generic Access Profile (GAP)
- [4] Blood Pressure Service Specification, Version 1.1 or later
- [5] Blood Pressure Service Specification, Version 1.1.1

3 Revision history and acknowledgments

Revision History

Publication Number	Revision Number	Date	Comments
0	1.0.0	2011-10-25	Adopted by the Bluetooth SIG Board of Directors
	1.0.1r00	2015-10-01	TSE 6611: Corrected conditionals in Table 2 for items 2/8, 2/9, 2/10, 2/11.
1	1.0.1	2015-12-22	Prepared for TCRL 2015-2 publication.
	1.0.2r00	2016-04-06	TSE 6996: Table numbering updated. Field deleted.
2	1.0.2	2016-07-13	Prepared for TCRL 2016-1 publication.
	1.0.3r00	2018-03-14	TSE 10482 (rating 3): Converted template. Added a Table 0. Revised reference for item 2/14. Added new item 2/15.
3	1.0.3	2018-06-27	Approved by BTI. Prepared for TCRL 2018-1 publication.
	1.0.4r00	2018-10-02	TSE 11004 (rating 1): Deleted reference “3.3” from Table 2, Item 3.
4	1.0.4	2018-11-21	Approved by BTI. Prepared for TCRL 2018-2 publication.
	1.0.5r00–r01	2019-04-02 – 2019-06-18	TSE 11364 (rating 1): Changed cross reference in Version table from note [1] to note [2].
5	1.0.5	2019-07-29	Approved by BTI. Prepared for TCRL 2019-1 publication.
	P6r00–r12	2018-06-10 – 2020-11-26	<p>Updated based on BLS v1.1 CR r03 specification.</p> <p>Updated based on BLS v1.1 CR r04 specification.</p> <p>Updated references based on BLS v1.1 CR r07.</p> <p>Updated ICS based on BLS v1.1 CR r08, introduced Table 2a, 2b and resolved BTI comments.</p> <p>Updated ICS based on BLS v1.1 PS.</p> <p>Removed from Table 2d the Operator <i>Equal to</i> as the Operator has been removed from the BLS v1.1 specification.</p> <p>Updated the ICS tables by removing the “Support” columns and for Tables 1 and 3 introduced the “Inter-Layer Dependency” column.</p> <p>Updated C.6 in Table 2, Service Requirements, to read “..otherwise Excluded”.</p>
6	p6	2020-12-22	<p>Set publication number for previous v1.0.5 to p5.</p> <p>Updated to incorporate test issue 16207. Template-related editorials. Approved by BTI on 2020-07-06.</p> <p>BLS v1.1 adopted by BoD on 2020-12-15. Prepared for publication.</p>

Publication Number	Revision Number	Date	Comments
	p7r00–r06	2021-02-22 – 2021-12-14	<p>TSE 16862 (rating 4): Added new ICS item 0b/1 for BLS v1.1.1 and new conditional C.3 in Table 0 for BLS v1.1. New Table 4 added for GAP inter-layer dependencies. Added new ICS item 2/25 and table (Table 2d) for E16461 - proper handling of changing supported features by characteristic indication. Added a reference for Blood Pressure Service Specification v1.1.1.</p> <p>TSE 17178 (rating 1): Updated Table 0 for the scheduled deprecation of BLP v1.0.</p> <p>TSE 17683 (rating 1): 2021 Deprecation and Withdrawal Adjustments. Updated Deprecation and Withdrawal dates in C.2 of Table 0 for BLS v1.0 after the 5 October 2021 Bluetooth SIG Board of Directors (BoD) approval to make adjustments to previously BoD-approved D&W dates. Minor editorials to align with current template conventions.</p> <p>Updated to latest template and made other editorials, including updating the copyright page to align with v2 of the DNMD. Consistency checker fixes.</p>
7	p7	2022-01-25	Approved by BTI on 2022-01-06. Prepared for TCRL 2021-2 publication.
	p8r00	2023-10-11	<p>TSE 23348 (rating 2): Resolved GATT inter-layer dependencies. In Table 0, marked Item 1 as Deprecated, updated the Status value for Items 1 and 2, marked C.1 and C.2 as no longer used, and updated C.3 and C.4. In Table 0b, updated C.1. In Table 2, updated the Status value for Items 16 and 25, updated C.1–C.5 and C.7–C.10, marked C.6 as no longer used, and deleted C.11. In Table 3, marked Item 1 as no longer used and updated the Capability and ILD values for Item 2.</p> <p>Made editorial edits to align the document with the latest ICS template, including updates to section and table titles and removal of “is supported” language in conditionals. Deleted draft revision history comments prior to p0.</p>
8	p8	2024-07-01	Approved by BTI on 2024-04-21. Prepared for TCRL 2024-1 publication.

Acknowledgments

Name	Company
Jörg Brakensiek	Bluetooth SIG, Inc.
Ismail Mohamud	Bluetooth SIG, Inc.
Alicia Courtney	Broadcom
Robert Hughes	Intel
Erik Moll	Koninklijke Philips N.V.
Tetsu Nishimura	Murata



Name	Company
Leif Aschehoug	Nordic Semiconductor
Magnus Sommansson	Qualcomm